



NOTES:

- 1 ALL RESISTANCES ARE IN OHMS.
- 2 ALL CAPACITIES IN MICROFARADS UNLESS OTHERWISE SPECIFIED.
- 3 S1 SHOWN IN 100 Ω POSITION.
- 4 S2 SHOWN IN DCV POSITION.
- 5 DISPLAYS D1 THRU D3 MUST BE REPLACED WITH SAME TYPE AND GRADE.
- 6 NETWORKS MUST BE REPLACED WITH SAME TYPE.
- 7 BATTERIES MUST BE INSTALLED IN UNIT IN ALL MODES OF OPERATION, AND POWER.
- 8 THIS SCHEMATIC SUBJECT TO CHANGE WITHOUT NOTICE.
- 9 C12 NOT USED.



Tavola C – Schema elettrico del multimetro digitale B + K PRECISION.